

info@lightscale.com 27 ORELAP #4112 OLCC #010-1003340D344


# **522 - Wizard of OG RCO 9.23.20**

Farmer's Friend Extracts 6451 NE Colwood Wy Portland, OR 97218 503-442-8653 Sample Type: Extracts Sample Date: 9/25/2020 Analysis Date: 9/30/2020 Report Date: 10/2/2020 Metrc Batch ID: 1A4010300016315000026969 Metrc Sample ID: 1A4010300016315000027027 Harvest/Process Date: 9/23/2020 Report ID: LS-201002-21 Sample Plan ID:SP-200924-8-A Sample Procedure: 160721\_LAB-SOP\_SampleCollection-v008

#### Potency

Potency Analysis Date: 10/1/2020 Potency Batch ID: CAN\_100120A Potency Method: JAOAC 2015.1

74.8%	Total THC
0.131%	Total CBD
Samples: PSP-TSF-BZX, FDS-NBH-RJR	



Analyte	Description	LOQ	RPD	Min.	Max.	Avg.	Unit: %
<b>Д9ТНС</b>	Delta-9 Tetrahydrocannabinol	0.16	1.14	74.4	75.2	74.8	
THCA	Tetrahydrocannabinolic acid	0.16	0.00	ND	ND	ND	
CBD	Cannabidiol	0.16	48.7	ND	0.263	0.131	•
CBDA	Cannabidiolic acid	0.16	0.00	ND	ND	ND	
∆8THC	Delta-8 Tetrahydrocannabinol*	0.16	0.00	ND	ND	ND	
THCV	Tetrahydrocannabivarin*	0.16	35.4	0.399	0.570	0.485	•
CBG	Cannabigerol*	0.16	42.7	0.518	0.799	0.659	•
CBGA	Cannabigerolic acid*	0.16	0.00	ND	<l0q< td=""><td><l0q< td=""><td></td></l0q<></td></l0q<>	<l0q< td=""><td></td></l0q<>	
CBC	Cannabichromene*	0.16	9.50	0.424	0.466	0.445	•
CBCA	Cannabichromenic acid*	0.16	0.00	ND	ND	ND	
CBN	Cannabinol	0.16	6.19	0.293	0.312	0.302	•
Total THC	Δ9THC + (THCA × 0.877)		1.14	74.4	75.2	74.8	
Total CBD	CBD + (CBDA × 0.877)		48.7	ND	0.263	0.131	•
Total			0.146	76.0	77.6	76.8	

#### Compliance

Pesticides	Within limits	Analysis Date: 10/1/2020	Pass ⊘
Solvents	Within limits	Analysis Date: 9/30/2020	Pass ⊘
Potency	Within limits	Analysis Date: 10/1/2020	Pass ⊘

Bryce Kidd, Ph.D. Lab Director

Aaron Troyer

Chief Science Officer



L S - 2 0 1 0 0 2 - 2 1

Lightscale Labs is accredited by ORELAP (Lab #4112) for analysis in compliance with OAR 333-064 and OAR 333-067. Results pertain to submitted samples only. Unless otherwise noted, samples were received in good condition and Quality Control samples met acceptance criteria. This Certificate shall not be reproduced except in full, without the written approval of Lightscale Labs. Results marked with an asterisk (\*) are not within scope of accreditation and for informational purposes only.



info@lightscale.com 27 ORELAP #4112 OLCC #010-1003340D344

# 522 - Wizard of OG RCO 9.23.20

Farmer's Friend Extracts 6451 NE Colwood Wy Portland, OR 97218 503-442-8653



#### Terpenes\* Sample Data

Sample Type: Extracts Sample Date: 9/25/2020 Analysis Date: 9/30/2020 Report Date: 10/2/2020 Metrc Batch ID:

1A4010300016315000026969 Metrc Sample ID: 1A4010300016315000027027 Harvest/Process Date: 9/23/2020 Report ID: LS-201002-21 Sample Plan ID:SP-200924-8-A Sample Procedure: 160721\_LAB-SOP\_SampleCollection-v008

Terpene Analysis Date: 9/30/2020 Terpene Batch ID: TRP\_093020A Method: JAOAC 2015.1 Unit: %

Analyte	Avg.	Notes
β-Caryophyllene	1.26%	
Humulene	0.377%	_
Linalool	0.315%	
Limonene	0.298%	
a-Terpineol	0.234%	-
β-Myrcene	0.137%	•
α-Bisabolol	0.129%	•
Selinadiene	0.0716%	•
β-Pinene	0.0404%	•
Sabinene	0.0349%	•
Terpinolene	0.0252%	•
Geraniol	0.0230%	•
α-Pinene	0.0123%	•
Guaiol	0.0110%	•
β-Ocimene	0.0102%	•
Fenchone	0.00874%	•
Eucalyptol	0.00557%	•
Camphene	0.00416%	•
a-Terpinene	0.00302%	•
γ-Terpinene	0.00289%	•
∆3-Carene	0.00220%	•
Nerol	0.00136%	•
Azulene	ND	
Borneol	ND	
Camphore	ND	
Caryophyllene Oxide	ND	
Cedrol	ND	
Cymene	ND	
Fenchol	ND	

Apolyto	A.v.a	Notoo
Analyte	Avg.	Notes
Geranyl Acetate	ND	
Isoborneol	ND	
Isopulegol	ND	
Pulegone	ND	
Sabinene Hydrate	ND	
Valencene	ND	
cis-Nerolidol	ND	
trans-Nerolidol	ND	
a-Cedrene	ND	
a-Ocimene	ND	
α-Phellandrene	ND	
β-Farnesene 1	ND	
β-Farnesene 2	ND	
γ-Terpineol	ND	
Total	3.01%	



Sample Type: Extracts

Sample Date: 9/25/2020

info@lightscale.com 7 ORELAP #4112 OLCC #010-1003340D344

# 522 - Wizard of OG RCO 9.23.20

Farmer's Friend Extracts 6451 NE Colwood Wy Portland, OR 97218 503-442-8653

Π

**Pesticides** 

Sample Data

#### Analysis Date: 9/30/2020 Report Date: 10/2/2020

Metrc Batch ID:

1A4010300016315000026969 Metrc Sample ID: 1A4010300016315000027027 Harvest/Process Date: 9/23/2020 Report ID: LS-201002-21 Sample Plan ID:SP-200924-8-A Sample Procedure: 160721\_LAB-SOP\_SampleCollection-v008

Pesticides Analysis Date: 10/1/2020 Pesticides Batch ID: PST\_100120A Method: EN 15662 Unit: ug/g (ppm) Pass ⊘

PSP-TSF-BZX FDS-NBH-RJR Limits Analyte LOQ Notes Status Abamectin ND ND 0.5 0.1 Pass Acenhate ND ND 0.4 0.1 Pass Acequinocyl ND ND 2.0 1.5 Pass Acetamiprid ND ND 0.2 0.1 Pass Aldicarb ND ND 0.4 0.1 Pass Azoxystrobin ND ND 0.2 0.1 Pass Bifenazate ND 0.2 ND 0.1 Pass Bifenthrin ND ND 0.2 0.1 Pass Boscalid ND ND 0.4 0.1 Pass Carbarvl ND ND A 2 0.1 Pass ND 0.2 Carbofuran ND 0.1 Pass Chlorantraniliprole ND ND 0.2 0.1 Pass Chlorfenapyr 0.1 ND ND 1.0 Pass Chlorpyrifos ND ND 0.2 0.1 Pass Clofentezine ND ND 0.2 0.1 Pass Cyfluthrin ND ND 1.0 0.5 Pass Cypermethrin ND ND 1.0 0.1 Pass Daminozide ND 1.0 0.5 ND Pass Diazinon ND ND 0.2 0.1 Pass Dichlorvos (DDVP) ND ND 1.0 0.5 Pass Dimethoate ND ND 0.2 0.1 Pass Ethoprophos ND ND 0.2 0.1 Pass Etofenprox ND ND 0.4 0.1 Pass ND Etoxazole ND A 2 0.1 Pass Fenoxycarb ND ND 0.2 0.1 Pass Fenpyroximate ND ND 0.4 0.1 Pass Fipronil ND ND 0.4 0.1 Pass Flonicamid ND ND 1.0 0.1 Pass Fludioxonil ND ND 0.4 0.1 Pass Hexythiazox ND ND 1.0 0.1 Pass Imazalil ND ND 0.2 0.1 Pass Imidacloprid ND ND 0.4 0.1 Pass Kresoxim-methyl ND ND 0.4 0.1 Pass Malathion ND ND 0.2 0.1 Pass

Analyte	PSP-TSF-BZX	FDS-NBH-RJR	Limits	LOQ	Notes	Status
Metalaxyl	ND	ND	0.2	0.1		Pass
Methiocarb	ND	ND	0.2	0.1		Pass
Methomyl	ND	ND	0.4	0.1		Pass
Methyl Parathion	ND	ND	0.2	0.2		Pass
MGK-264	ND	ND	0.2	0.2		Pass
Myclobutanil	ND	ND	0.2	0.1		Pass
Naled	ND	ND	0.5	0.2		Pass
Oxamyl	ND	ND	1.0	0.1		Pass
Paclobutrazol	ND	ND	0.4	0.1		Pass
Permethrins	ND	ND	0.2	0.1		Pass
Phosmet	ND	ND	0.2	0.1		Pass
Piperonyl Butoxide	ND	ND	2.0	1.0		Pass
Prallethrin	ND	ND	0.2	0.1		Pass
Propiconazole	ND	ND	0.4	0.1		Pass
Propoxur	ND	ND	0.2	0.1		Pass
Pyrethrins	ND	ND	1.0	0.5		Pass
Pyridaben	ND	ND	0.2	0.1		Pass
Spinosad	ND	ND	0.2	0.1		Pass
Spiromesifen	ND	ND	0.2	0.1		Pass
Spirotetramat	ND	ND	0.2	0.1		Pass
Spiroxamine	ND	ND	0.4	0.1		Pass
Tebuconazole	ND	ND	0.4	0.1		Pass
Thiacloprid	ND	ND	0.2	0.1		Pass
Thiamethoxam	ND	ND	0.2	0.1		Pass
Trifloxystrobin	ND	ND	0.2	0.1		Pass



info@lightscale.com 7 ORELAP #4112 OLCC #010-1003340D344

# 522 - Wizard of OG RCO 9.23.20

Farmer's Friend Extracts 6451 NE Colwood Wy Portland, OR 97218 503-442-8653

IJ

Sample Type: Extracts Sample Date: 9/25/2020 Analysis Date: 9/30/2020 Report Date: 10/2/2020 Metrc Batch ID: 1A4010300016315000026969 Metrc Sample ID: 1A4010300016315000027027 Harvest/Process Date: 9/23/2020 Report ID: LS-201002-21 Sample Plan ID:SP-200924-8-A Sample Procedure: 160721\_LAB-SOP\_SampleCollection-v008

Pesticides QC Analysis Date: 10/1/2020 Pesticides QC Batch ID: PST\_100120A Method: EN 15662 Unit: µg/g (ppm)

#### Pesticides Quality Control Data

Analyte	Blank	LOQ	LCS	LCS Spike	LCS Rec (%)	Limits (%)	Notes
Abamectin	ND	0.1	0.776	1.00	77.6	50 - 150	
Acephate	ND	0.1	0.907	1.00	90.7	50 - 150	
Acequinocyl	ND	1.5	0.755	1.00	75.5	50 - 150	
Acetamiprid	ND	0.1	0.992	1.00	99.2	50 - 150	
Aldicarb	ND	0.1	0.951	1.00	95.1	50 - 150	
Azoxystrobin	ND	0.1	1.04	1.00	104	50 - 150	
Bifenazate	ND	0.1	1.12	1.00	112	50 - 150	
Bifenthrin	ND	0.1	0.868	1.00	86.8	50 - 150	
Boscalid	ND	0.1	0.977	1.00	97.7	50 - 150	
Carbaryl	ND	0.1	0.990	1.00	99.0	50 - 150	
Carbofuran	ND	0.1	0.974	1.00	97.4	50 - 150	
Chlorantraniliprole	ND	0.1	0.900	1.00	90.0	50 - 150	
Chlorfenapyr	ND	0.1	0.887	1.00	88.7	50 - 150	
Chlorpyrifos	ND	0.1	0.962	1.00	96.2	50 - 150	
Clofentezine	ND	0.1	0.653	1.00	65.3	50 - 150	
Cyfluthrin	ND	0.5	0.889	1.00	88.9	50 - 150	
Cypermethrin	ND	0.1	1.00	1.00	100	50 - 150	
Daminozide	ND	0.5	0.322	1.00	32.2	10 - 150	
Diazinon	ND	0.1	1.02	1.00	102	50 - 150	
Dichlorvos (DDVP)	ND	0.5	0.961	1.00	96.1	50 - 150	
Dimethoate	ND	0.1	0.917	1.00	91.7	50 - 150	
Ethoprophos	ND	0.1	1.03	1.00	103	50 - 150	
Etofenprox	ND	0.1	0.897	1.00	89.7	50 - 150	
Etoxazole	ND	0.1	0.982	1.00	98.2	50 - 150	
Fenoxycarb	ND	0.1	1.04	1.00	104	50 - 150	
Fenpyroximate	ND	0.1	0.846	1.00	84.6	50 - 150	
Fipronil	ND	0.1	1.02	1.00	102	50 - 150	
Flonicamid	ND	0.1	0.966	1.00	96.6	50 - 150	
Fludioxonil	ND	0.1	2.58	1.00	258	50 - 150	HB
Hexythiazox	ND	0.1	0.871	1.00	87.1	50 - 150	
Imazalil	ND	0.1	1.10	1.00	110	50 - 150	
Imidacloprid	ND	0.1	0.959	1.00	95.9	50 - 150	
Kresoxim-methyl	ND	0.1	1.06	1.00	106	50 - 150	
Malathion	ND	0.1	1.05	1.00	105	50 - 150	

Analyte	Blank	LOQ	LCS	LCS Spike	LCS Rec (%)	Limits (%)	Notes
Metalaxyl	ND	0.1	0.993	1.00	99.3	50 - 150	
Methiocarb	ND	0.1	0.980	1.00	98.0	50 - 150	
Methomyl	ND	0.1	0.897	1.00	89.7	50 - 150	
Methyl Parathion	ND	0.2	ND	1.00	0.00	30 - 150	ND
MGK-264	ND	0.2	0.206	0.600	34.4	50 - 150	LR
Myclobutanil	ND	0.1	1.04	1.00	104	50 - 150	
Naled	ND	0.2	0.878	1.00	87.8	50 - 150	
Oxamyl	ND	0.1	0.924	1.00	92.4	50 - 150	
Paclobutrazol	ND	0.1	0.882	1.00	88.2	50 - 150	
Permethrins	ND	0.1	1.02	1.00	102	50 - 150	
Phosmet	ND	0.1	1.13	1.00	113	50 - 150	
Piperonyl Butoxide	ND	1.0	0.844	1.00	84.4	50 - 150	
Prallethrin	ND	0.1	1.20	1.00	120	50 - 150	
Propiconazole	ND	0.1	1.52	1.00	152	50 - 150	ME
Propoxur	ND	0.1	0.946	1.00	94.6	50 - 150	
Pyrethrins	ND	0.5	0.786	1.00	78.6	50 - 150	
Pyridaben	ND	0.1	0.952	1.00	95.2	50 - 150	
Spinosad	ND	0.1	0.782	1.00	78.2	50 - 150	
Spiromesifen	ND	0.1	1.03	1.00	103	50 - 150	
Spirotetramat	ND	0.1	0.985	1.00	98.5	50 - 150	
Spiroxamine	ND	0.1	0.808	1.00	80.8	50 - 150	
Tebuconazole	ND	0.1	1.35	1.00	135	50 - 150	
Thiacloprid	ND	0.1	1.07	1.00	107	50 - 150	
Thiamethoxam	ND	0.1	1.03	1.00	103	50 - 150	
Trifloxystrobin	ND	0.1	0.912	1.00	91.2	50 - 150	

Lightscale Labs is accredited by ORELAP (Lab #4112) for analysis in compliance with OAR 333-064 and OAR 333-067. Results pertain to submitted samples only. Unless otherwise noted, samples were received in good condition and Quality Control samples met acceptance criteria. This Certificate shall not be reproduced except in full, without the written approval of Lightscale Labs. Results marked with an asterisk (\*) are not within scope of accreditation and for informational purposes only.



info@lightscale.com 7 ORELAP #4112 OLCC #010-1003340D344

# 522 - Wizard of OG RCO 9.23.20

Farmer's Friend Extracts 6451 NE Colwood Wy Portland, OR 97218 503-442-8653 Sample Type: Extracts Sample Date: 9/25/2020 Analysis Date: 9/30/2020 Report Date: 10/2/2020



1A4010300016315000026969 Metrc Sample ID: 1A4010300016315000027027

Metrc Batch ID:

Harvest/Process Date: 9/23/2020 Report ID: LS-201002-21 Sample Plan ID:SP-200924-8-A Sample Procedure: 160721\_LAB-SOP\_SampleCollection-v008

Pass ⊘

Solvents Analysis Date: 9/30/2020 Solvents Batch ID: RES\_092920B Method: EPA 5021A Unit: µg/g (ppm)

Analyte	PSP-TSF-BZX	FDS-NBH-RJR	RPD (%)	Limits	LOQ	Notes	Status
1,4-Dioxane	ND	ND	0.00	380.0	50.0		Pass
2-Butanol	ND	ND	0.00	5000.0	250.0		Pass
2-Ethoxyethanol	ND	ND	0.00	160.0	50.0		Pass
Acetone	ND	ND	0.00	5000.0	250.0		Pass
Acetonitrile	ND	ND	0.00	410.0	50.0		Pass
Benzene	ND	ND	0.00	2.0	2.0		Pass
Butanes	ND	ND	0.00	5000.0	250.0		Pass
Cumene	ND	ND	0.00	70.0	50.0		Pass
Cyclohexane	ND	ND	0.00	3880.0	50.0		Pass
Ethyl Acetate	ND	ND	0.00	5000.0	250.0		Pass
Ethyl Ether	ND	ND	0.00	5000.0	250.0		Pass
Ethylene Glycol	ND	ND	0.00	620.0	250.0		Pass
Ethylene Oxide	ND	ND	0.00	50.0	50.0		Pass
Heptane	ND	ND	0.00	5000.0	250.0		Pass
Hexanes	ND	ND	0.00	290.0	50.0		Pass
Isopropanol (2-Propanol)	ND	ND	0.00	5000.0	50.0		Pass
Isopropyl Acetate	ND	ND	0.00	5000.0	250.0		Pass
Methanol	ND	ND	0.00	3000.0	250.0		Pass
Dichloromethane	ND	ND	0.00	600.0	50.0		Pass
Pentanes	ND	ND	0.00	5000.0	250.0		Pass
Propane	ND	ND	0.00	5000.0	250.0		Pass
Tetrahydrofuran	ND	ND	0.00	720.0	50.0		Pass
Toluene	ND	ND	0.00	890.0	50.0		Pass
Xylenes	ND	ND	0.00	2170.0	50.0		Pass



info@lightscale.com 7 ORELAP #4112 OLCC #010-1003340D344

# 522 - Wizard of OG RCO 9.23.20

Farmer's Friend Extracts 6451 NE Colwood Wy Portland, OR 97218 503-442-8653

0

**Residual Solvents** 

**Quality Control Data** 

Sample Type: Extracts Sample Date: 9/25/2020 Analysis Date: 9/30/2020 Report Date: 10/2/2020 Metrc Batch ID:

1A4010300016315000026969 Metrc Sample ID: 1A4010300016315000027027 Harvest/Process Date: 9/23/2020 Report ID: LS-201002-21 Sample Plan ID:SP-200924-8-A Sample Procedure: 160721\_LAB-SOP\_SampleCollection-v008

Solvents QC Analysis Date: 9/30/2020 Solvents QC Batch ID: RES\_092920B Method: EPA 5021A Unit: µg/g (ppm)

Analyte	Blank	LOQ	LCS	LCS Spike	LCS Rec (%)	Limits (%)		Notes
1,4-Dioxane	ND	50.0	1110	1000	111	70 - 130		
2-Butanol	ND	250.0	989	1000	98.9	70 - 130		
2-Ethoxyethanol	ND	50.0	1300	1000	130	70 - 130		
Acetone	ND	250.0	1210	1000	121	70 - 130		
Acetonitrile	ND	50.0	1290	1000	129	70 - 130		
Benzene	ND	2.0	19.9	20.0	99.5	70 - 130		
Butanes	ND	250.0	2150	2000	108	70 - 130		
Cumene	ND	50.0	1050	1000	105	70 - 130		
Cyclohexane	ND	50.0	1150	1000	115	70 - 130		
Ethyl Acetate	ND	250.0	1020	1000	102	70 - 130		
Ethyl Ether	ND	250.0	1310	1000	131	70 - 130	ME	
Ethylene Glycol	ND	250.0	1370	1000	137	70 - 130	ME	
Ethylene Oxide	ND	50.0	2.11	1000	106.5	70 - 130		
Heptane	ND	250.0	1010	1000	101	70 - 130		
Hexanes	ND	50.0	5600	5000	112	70 - 130		
Isopropanol (2-Propanol)	ND	50.0	1010	1000	101	70 - 130		
Isopropyl Acetate	ND	250.0	1010	1000	101	70 - 130		
Methanol	ND	250.0	1220	1000	122	70 - 130		
Dichloromethane	ND	50.0	1020	1000	102	70 - 130		
Pentanes	ND	250.0	3260	3000	109	70 - 130		
Propane	ND	250.0	1220	1000	122	70 - 130		
Tetrahydrofuran	ND	50.0	1050	1000	105	70 - 130		
Toluene	ND	50.0	1120	1000	112	70 - 130		
Xylenes	ND	50.0	4770	4000	119	70 - 130		



info@lightscale.com 27 ORELAP #4112 OLCC #010-1003340D344

#### 522 - Wizard of OG RCO 9.23.20

Farmer's Friend Extracts 6451 NE Colwood Wy Portland, OR 97218 503-442-8653 Sample Type: Extracts Sample Date: 9/25/2020 Analysis Date: 9/30/2020 Report Date: 10/2/2020 Metrc Batch ID: 1A4010300016315000026969 Metrc Sample ID: 1A4010300016315000027027 Harvest/Process Date: 9/23/2020 Report ID: LS-201002-21 Sample Plan ID:SP-200924-8-A Sample Procedure: 160721\_LAB-SOP\_SampleCollection-v008

#### **Qualifier Flag Descriptions**

- J Reported result is an estimate the value is less than the minimum calibration level but greater than the estimated detection limit (EDL)
- U The analyte was not detected in the sample at the estimated detection limit (EDL)
- E Exceeds calibration range
- D Dilution data result was obtained from the analysis of a dilution
- B Analyte found in sample and associated blank
- C Co-eluting compound
- R Relative Percent Difference (RPD) outside control limits
- NR Analyte not reported because of problems in sample preparation or analysis
- ND Non-Detect
- X Results from reinjection/repeat/re-column data
- EMC Estimated maximum possible concentration indicates that a peak is detected but did not meet the method required criteria
- M Manual integration
- PS Peaks split
- HB Control acceptance criteria are exceeded high and the associated sample is below the detection limit
- LB Control acceptance criteria are exceeded low and the associated sample exceeds the regulatory limit
- ME Marginal Exceedance
- LR Low Recovery Analyte
- LOQ Limit of Quantitation